





g dose.


. See below for

ROM LIGHT,

ch containing dles.

Icohol, NF; and

nt and release





7. Attach the safety needle to the syringe tip (Luer-Lok). Twist needle until firmly seated (see Figure 3). Confirm that the needle is locked to the Luer connector before moving or tilting the syringe out of the vertical plane to avoid spillage of syringe contents.

For Administration:

- 8. Pull shield straight off needle to avoid damaging needle point.
- 9. Remove needle sheath.
- 10. Expel excess gas from the syringe (a small gas bubble may remain).
- 11. Administer intramuscularly slowly (1-2 minutes/injection) into the buttock (gluteal area). For user convenience, the needle 'bevel up' position is orientated to the lever arm, as shown in Figure 4.
- 12. After injection, immediately activate the lever arm to deploy the needle shielding by applying a single-finger stroke to the activation assisted lever arm to push the lever arm completely forward. Listen for a click. Confirm that the needle shielding has completely covered the needle (see Figure 5). NOTE: Activate away from self and others.
- 13. Discard the empty single use syringe into an approved sharps collector in accordance with applicable regulations and institutional policy.
- 14. Repeat steps 1 through 13 for second syringe.

SafetyGlide $^{\text{TM}}$ is a trademark of Becton Dickinson and Company. FASLODEX is a trademark of the AstraZeneca group of companies. © AstraZeneca 2016

Distributed by: AstraZeneca Pharmaceuticals LP, Wilmington, DE 19850 Manufactured for: AstraZeneca UK Ltd., Macclesfield, England By: Vetter Pharma-Fertigung GmbH & Co. KG, Ravensburg, Germany Product of Germany





Figure 4



Figure 5





Both syringes must be administered to receive the 500 mg dose Omg/ Û





inges must be administered e the 500 mg dose





This is a representation of an electronic record that was signed electronically and this page is the manifestation of the electronic signature.	-
/s/	-
AMNA IBRAHIM 01/24/2017	

