

K031780
Page 1 of 2

JUL 30 2003

VERTE-STACK™ Spinal System
510(k) Summary
June 2003

- I. **Company:** Medtronic Sofamor Danek
1800 Pyramid Place
Memphis, TN 38132
(901) 396-3133
- II. **Proprietary Trade Name:** VERTE-STACK™ Spinal System
- III. **Classification Name:** Spinal Intervertebral Body Fixation Orthosis (21 CFR 888.3060)
- III. **Product Description**

The VERTE-STACK™ device, is a stackable PEEK spacer, which inserts between vertebral bodies in the anterior thoracic and lumbar spine. The device is intended for vertebral body replacement to aid in the surgical correction and stabilization of the spine. The construct is not intended to be employed as a stand-alone device. The VERTE-STACK™ device is fabricated and manufactured from POLYETHERETHERKETONE (PEEK OPTIMA LT) as described by ASTM F2026. The tantalum marker used for this product is made to the voluntary standard of ASTM F-560.

The design of the VERTE-STACK™ device includes a variety of stackable components of different sizes and heights. The stackable components are designed to suit the individual patient pathology.

The VERTE-STACK™ device may be used individually, or stacked together in order to accommodate the individual anatomical requirements of the vertebral space created by the corpectomy.

The VERTE-STACK™ Spinal System must be used with additional anterior and/or posterior spinal instrumentation to augment stability.

The purpose of this submission is to add modified components to the VERTE-STACK™ Spinal System.

IV. **Indications**

The VERTE-STACK™ Spinal System is a vertebral body replacement device intended for use in the thoracolumbar spine (T1-L5) to replace a collapsed, damaged, or unstable vertebral body due to tumor or trauma (i.e., fracture). The VERTE-STACK™ Spinal System is to be used with supplemental fixation. Specifically, the VERTE-STACK™



000109

K031786

Page 2 of 2

device is to be used with the Medtronic Sofamor Danek ZPLATE II Anterior Fixation System, the Titanium DYNALOK™ CLASSIC Spinal System, the VANTAGE™ Anterior Fixation System, the Titanium TSRH® Spinal System, the Titanium CD HORIZON® Spinal System, the Titanium GDLH® Spinal System or their successors. Additionally, the VERTE-STACK™ device is intended to be used with bone graft.

V. Substantial Equivalence

Documentation was provided which demonstrated the VERTE-STACK™ Spinal System to be substantially equivalent to the previously cleared VERTE-STACK™ Spinal System (K023570, SE 11/19/02 and K021791, SE 08/26/02).

000110



Food and Drug Administration
9200 Corporate Boulevard
Rockville MD 20850

JUL 30 2003

Richard W. Treharne, Ph.D.
Sr. Vice President Regulatory Affairs
Medtronic Sofamor Danek
1800 Pyramid Place
Memphis, TN 38132

Re: K031780
Trade/Device Name: VERTE-STACK™ Spinal System
Regulation Number: 21 CFR 888.3060
Regulation Name: Spinal intervertebral body fixation orthosis
Regulatory Class: II
Product Code: MQP
Dated: June 27, 2003
Received: June 30, 2003

Dear Dr. Treharne:

We have reviewed your Section 510(k) premarket notification of intent to market the device referenced above and have determined the device is substantially equivalent (for the indications for use stated in the enclosure) to legally marketed predicate devices marketed in interstate commerce prior to May 28, 1976, the enactment date of the Medical Device Amendments, or to devices that have been reclassified in accordance with the provisions of the Federal Food, Drug, and Cosmetic Act (Act) that do not require approval of a premarket approval application (PMA). You may, therefore, market the device, subject to the general controls provisions of the Act. The general controls provisions of the Act include requirements for annual registration, listing of devices, good manufacturing practice, labeling, and prohibitions against misbranding and adulteration.

If your device is classified (see above) into either class II (Special Controls) or class III (PMA), it may be subject to such additional controls. Existing major regulations affecting your device can be found in the Code of Federal Regulations, Title 21, Parts 800 to 898. In addition, FDA may publish further announcements concerning your device in the Federal Register.


Please be advised that FDA's issuance of a substantial equivalence determination does not mean that FDA has made a determination that your device complies with other requirements of the Act or any Federal statutes and regulations administered by other Federal agencies. You must comply with all the Act's requirements, including, but not limited to: registration and listing (21 CFR Part 807); labeling (21 CFR Part 801); good manufacturing practice requirements as set forth in the quality systems (QS) regulation (21 CFR Part 820); and if applicable, the electronic product radiation control provisions (Sections 531-542 of the Act); 21 CFR 1000-1050.

Page 2 - Richard W. Treharne, Ph.D.

This letter will allow you to begin marketing your device as described in your Section 510(k) premarket notification. The FDA finding of substantial equivalence of your device to a legally marketed predicate device results in a classification for your device and thus, permits your device to proceed to the market.

If you desire specific advice for your device on our labeling regulation (21 CFR Part 801), please contact the Office of Compliance at (301) 594-4659. Also, please note the regulation entitled, "Misbranding by reference to premarket notification" (21 CFR Part 807.97). You may obtain other general information on your responsibilities under the Act from the Division of Small Manufacturers, International and Consumer Assistance at its toll-free number (800) 638-2041 or (301) 443-6597 or at its Internet address <http://www.fda.gov/cdrh/dsma/dsmamain.html>

Sincerely yours,

for Celia M. Witten, Ph.D., M.D.

Director
Division of General, Restorative
and Neurological Devices
Office of Device Evaluation
Center for Devices and
Radiological Health

Enclosure

510(k) Number (if known): K031780

Device Name: VERTE-STACK™ Spinal System

Indications for Use:

The VERTE-STACK™ Spinal System is a vertebral body replacement device intended for use in the thoracolumbar spine (T1-L5) to replace a collapsed, damaged, or unstable vertebral body due to tumor or trauma (i.e., fracture). The VERTE-STACK™ Spinal System is to be used with supplemental fixation. Specifically, the VERTE-STACK™ device is to be used with the Medtronic Sofamor Danek ZPLATE II Anterior Fixation System, Titanium DYNALOK™ CLASSIC Spinal System, the VANTAGE™ Anterior Fixation System, Titanium TSRH® Spinal System, Titanium CD HORIZON® Spinal System, the Titanium GDLH® Spinal System, or their successors. Additionally, the VERTE-STACK™ device is intended to be used with bone graft.


(PLEASE DO NOT WRITE BELOW THIS LINE—CONTINUE ON ANOTHER PAGE IF NEEDED)

Concurrence of CDRH, Office of Evaluation (ODE)

Prescription Use _____
(Per 21 CFR 801.109)

OR

Over-The-Counter Use _____
(Optional 1-2-96)



(Division Sign-Off)
Division of General, Restorative
and Neurological Devices

510(k) Number K031780

000158

Explore Litigation Insights

Docket Alarm provides insights to develop a more informed litigation strategy and the peace of mind of knowing you're on top of things.

Real-Time Litigation Alerts



Keep your litigation team up-to-date with **real-time alerts** and advanced team management tools built for the enterprise, all while greatly reducing PACER spend.

Our comprehensive service means we can handle Federal, State, and Administrative courts across the country.

Advanced Docket Research



With over 230 million records, Docket Alarm's cloud-native docket research platform finds what other services can't. Coverage includes Federal, State, plus PTAB, TTAB, ITC and NLRB decisions, all in one place.

Identify arguments that have been successful in the past with full text, pinpoint searching. Link to case law cited within any court document via Fastcase.

Analytics At Your Fingertips



Learn what happened the last time a particular judge, opposing counsel or company faced cases similar to yours.

Advanced out-of-the-box PTAB and TTAB analytics are always at your fingertips.

API

Docket Alarm offers a powerful API (application programming interface) to developers that want to integrate case filings into their apps.

LAW FIRMS

Build custom dashboards for your attorneys and clients with live data direct from the court.

Automate many repetitive legal tasks like conflict checks, document management, and marketing.

FINANCIAL INSTITUTIONS

Litigation and bankruptcy checks for companies and debtors.

E-DISCOVERY AND LEGAL VENDORS

Sync your system to PACER to automate legal marketing.