Journal of Chromatography, 260 (1983) 419-427 Elsevier Science Publishers B.V., Amsterdam — Printed in the Netherlands

CHROM. 15,668

COMPARISON OF SEVERAL HIGH-PERFORMANCE LIQUID CHROMA-TOGRAPHY TECHNIQUES FOR THE SEPARATION OF OLIGODEOXY-NUCLEOTIDES ACCORDING TO THEIR CHAIN LENGTHS

WOLFGANG HAUPT and ALFRED PINGOUD* Abteilung Biophysikalische Chemie, Zentrum Biochemie, Medizinische Hochschule Hannover, Konstanty-Gutschow-Strasse, D-3000 Hannover 61 (G.F.R.) (Received December 10th, 1982)

SUMMARY

The use of reversed-phase, reversed-phase-ion-pair and anion-exchange highperformance liquid chromatography (HPLC) was investigated for the analytical and preparative separation of oligodeoxynucleotides according to their chain length. The data obtained with homooligonucleotides and oligonucleotides of defined sequence show that reversed-phase-ion-pair and anion-exchange, but not reversed-phase, HPLC can be used reliably to separate oligodeoxynucleotides according to their chain length, largely irrespective of their base composition. The chain length limits for complete separation within 30 min by reversed-phase, reversed-phase-ion-pair and anion-exchange HPLC are approximately 10, 15 and 20 nucleotides, respectively.

INTRODUCTION

DOCKE.

High-performance liquid chromatography (HPLC), because of its high resolution, reproducibility and ease of operation, has become an indispensible method for the separation of oligodeoxynucleotides, for both analytical and preparative purposes. A variety of different techniques have been developed, based on reversed-phase HPLC¹⁻³, reversed-phase-ion-pair HPLC^{4,5}, stronganion-exchange HPLC⁶⁻⁹ and size exclusion HPLC¹⁰, and have been used so far mainly for the separation of short oligonucleotides. The separation of oligodeoxynucleotides by these techniques is determined by the base composition and chain length of the oligodeoxynucleotides to be separated, which in turn determine their polarity and size. These properties, albeit to different extents in the various chromatographic systems, control the degree of retention of a particular oligodeoxynucleotide. The retention mechanisms involved are not fully understood.

Reversed-phase HPLC of oligodeoxynucleotides is dominated by hydrophobic interactions between the solute and the bonded phase¹; it is more sensitive to the base composition than reversed-phase-ion-pair, anion-exchange and size exclusion HPLC. Nevertheless the retention sequence of oligonucleotides with similar base composition is controlled by the charge of the oligonucleotide.

In reversed-phase-ion-pair HPLC, negative charges on the oligodeoxynucleo-

0021-9673/83/\$03.00 © 1983 Elsevier Science Publishers B.V.

CUREVAC EX2029 Page 1

Find authenticated court documents without watermarks at docketalarm.com.

tide are neutralized by positively charged alkylammonium ions. It is not clear, however, whether this process occurs in the mobile phase¹¹ or in the stationary phase¹², leading to dynamic ion exchange. The retention is determined mainly by the charge of the oligodeoxynucleotide; in addition to these dominating electrostatic interactions, hydrophobic interactions between the oligodeoxynucleotides and the reversed phase also play a role in the chromatographic process⁵.

The separation of oligodeoxynucleotides by anion-exchange HPLC is dependent on differences in charge; as currently used anion exchangers are based on a silica gel matrix, in which the tertiary ammonium moiety is connected to the resin via an aliphatic spacer, there are hydrophobic interactions in addition to the ionic interactions. The inclusion of non-polar solvents in the mobile phase diminishes the effect of the reversed phase on the chromatography^{13,14}.

Recently, silica gel bonded polyol phases have become commercially available for the separation of oligodeoxynucleotides by size exclusion HPLC. The separation can be carried out with a variety of mobile phases. The resolution depends on the mobile phase, indicating that size is not the only determinant for the chromatography. For small and medium sized oligodeoxynucleotides the resolution is not as good as with the above-mentioned chromatographic systems¹⁰.

One characteristic feature of a particular oligodeoxynucleotide, and one that can be used conveniently for the purpose of analysis or preparative purification, is its size. It is therefore desirable to have chromatographic systems that allow separations according to size, *i.e.*, chain length. A separation in such a chromatographic system ideally should be unaffected by base composition and should allow separation with high resolution over a wide range of chain lengths. For this purpose we investigated the suitability of two standard and one new chromatographic system for chain length-dependent separations of oligodeoxynucleotides, namely (1) reversed-phase HPLC on Zorbax ODS columns, (2) reversed-phase-ion-pair HPLC on LiChrosorb RP-8 columns and (3) anion-exchange HPLC on Partisil SAX columns. We used commercial homooligodeoxynucleotides of specified chain lengths as well as chemically or enzymatically synthesized oligodeoxynucleotides of defined sequence to study the limits of resolution of these chromatographic systems.

The results show that both reversed-phase-ion-pair HPLC and anion-exchange HPLC, but not reversed-phase HPLC, can be used reliably to separate oligodeoxynucleotides according to their chain length, with only slight interference by base composition. Whereas, reversed-phase-ion-pair HPLC has a cut-off for an acceptable resolution at a chain length of around 15–20 nucleotides, anion-exchange HPLC allows even longer oligodcoxynucleotides to be separated.

EXPERIMENTAL

Instrumentation

DOCKE.

Two chromatographs were used: (1) a DuPont 850 liquid chromatograph, consisting of a gradient controller, a three-head pump, a temperature-controlled column compartment, a UV spectrophotometer and a Rheodyne 7125 injection valve, and (2) a Pye Unicam liquid chromatograph, consisting of an LC-XP gradient programmer, an Altex Model 100A pump, an LC-UV detector and a Rheodyne 7125 injection valve, to which a Merck LiChrocart-Autofix pneumatic mounting device was attached.

> CUREVAC EX2029 Page 2

420

HPLC OF OLIGODEOXYNUCLEOTIDES

Chemicals and oligonucleotides

All homodeoxyoligonucleotides were obtained from P.L. Biochemicals (St. Goar, G.F.R.). d(GGAATTCC) and d(CGAATTCG) were kindly given by Dr. H. Bloecker and Dr. R. Frank (GBF, Stöckheim, G.F.R.) and by Dr. M. Zabeau (EMBL, Heidelberg, G.F.R.), respectively. Tetrabutylammonium hydrogen sulphate (TBAHSO₄), triethylamine and formamide were products of Fluka (Neu-Ulm, G.F.R.) and were used without further purification. HPLC-grade acetonitrile was a product of Burdick & Jackson Labs. (Muskegon, MI, U.S.A.) and was provided by Macherey, Nagel & Co. (Düren, G.F.R.). T4 DNA ligase was obtained from BRL (Neu-Isenberg, G.F.R.). All other chemicals used were of pro analysi grade from Merck (Darmstadt, G.F.R.).

Columns and chromatographic conditions

All eluents were filtered through a Whatman GF/F glass-microfibre-filter (Vetter, Wiesloch, G.F.R.) and degassed prior to chromatography.

For reversed-phase separations Zorbax ODS columns (25 cm \times 4.6 mm I.D., 5 μ m) from DuPont (Bad Nauheim, G.F.R.) were used. The triethylammonium acetate (TEAA) buffer was prepared as described elsewhere¹. Solvent A was 0.1 \dot{M} TEAA (pH 7.0)–1% acetonitrile and solvent B was 0.1 *M* TEAA (pH 7.0)–50% acetonitrile. If not stated otherwise, a linear gradient from 15 to 30% B in 30 min was applied. The flow-rate was 1.0 ml/min. Chromatography was carried out at ambient temperature.

Reversed-phase-ion-pair separations were performed on Hibar LiChrocart cartridges (25 cm \times 4 mm I.D.) packed with LiChrosorb RP-8, 10 μ m (Merck). Solvent A was 50 mM potassium phosphate (pH 5.9)-2 mM TBAHSO₄ and solvent B was 50 mM potassium phosphate (pH 6.5)-2 mM TBAHSO₄-60% acetonitrile. The standard linear gradient used was from 15 to 80% B in 60 min. The flow-rate was 1.0 ml/min. Chromatography was carried out at ambient temperature.

Whatman Partisil 10 SAX columns $(25 \times 4.6 \text{ mm I.D.}, 10 \,\mu\text{m})$ obtained from IC-Chemikalien (Munich, G.F.R.) were used for anion-exchange HPLC. Solvent A was 1 mM potassium phosphate (pH 6.3) in formamide-water (6:4) and solvent B was 0.3 M potassium phosphate (pH 6.3) in formamide-water (6:4). The linear gradient used was from 0 to 80% B in 60 min. The flow-rate was 1.0 ml/min. Chromatography was carried out at 45°C.

RESULTS

DOCKE⁻

Figs. 1 and 2 show chromatograms of mixtures of oligo-(dA)s and oligo-(dT)s separated by reversed-phase chromatography. Oligonucleotides of different chain length are well resolved within the oligo-(dA) and oligo-(dT) series up to a chain length of at least ten nucleotides under the conditions given.

Oligo-(dT)s are considerably more retarded than oligo-(dA)s of the same chain length. Thus, for example, $p(dA)_{10}$ is eluted before $p(dT)_6$. This demonstrates that the base composition of the oligonucleotide has a considerable effect on the retention of the oligonucleotides. This effect, of course, can be of advantage when oligomers of similar chain length or even sequence isomers have to be separated. This is illustrated in Fig. 3, where a separation of d(GGAATTCC) from its sequence isomer

CUREVAC EX2029 Page 3



422

Fig. 1. Reversed-phase HPLC of oligo-(dA)s. Column: Zorbax ODS (25 cm \times 3.6 mm I.D.). Eluents: A, 0.1 *M* TEAA (pH) 7.0–1% acetonitrile; B, 0.1 *M* TEAA (pH 7.0)–50% acetonitrile; linear gradient from 15 to 30% B in 30 min at ambient temperature; flow-rate, 1.0 ml/min. Unidentified peaks are due to contaminants in the individual samples.

Fig. 2. Reversed-phase HPLC of oligo-(dT)s. Chromatographic conditions as in Fig. 1.

d(CGAATTCG) is shown. In fact, our analysis of commercially available oligo-(dT)s and in particular of oligo-(dA)s has shown that these preparations contain considerable amounts of impurities (*cf.*, Figs. 1 and 2), which are not detected by polyacrylamide gel electrophoresis under denaturing conditions, by homochromatography after labelling the 5'-end with [³²P]phosphate or by anion-exchange HPLC (see below), suggesting that they consist of chemical derivatives of these oligonucleotides with the same chain length or charge.

Figs. 4 and 5 show separations of mixtures of oligo-(dA)s and oligo-(dT)s in a reversed-phase-ion-pair system. The individual oligonucleotides are extremely well resolved up to a chain length of at least 16 nucleotides, the peaks being sharper and more symmetrical than in the reversed-phase system (compare Figs. 4 and 1 or Figs. 5 and 2). Oligo-(dA)s are slightly more retained than oligo-(dT)s, *e.g.*, $p(dA)_8$ is eluted after $p(dT)_{10}$. The influence of base composition is less pronounced in the reversed-phase system. Accordingly, reversed-phase-ion-pair HPLC is not as useful as reversed-phase HPLC for the separation of oligonucleotides of different base composition but similar chain length. This is also apparent from Fig. 6, which shows the

> CUREVAC EX2029 Page 4

Find authenticated court documents without watermarks at docketalarm.com.



Fig. 3. Reversed-phase HPLC of the sequence isomers d(GGAATTCC) and d(CGAATTCG). Column: Zorbax ODS (25 cm × 4.6 mm I.D.). Eluents: A, 0.1 M TEEA (pH 7.0)-1% acetonitrile; B, 0.1 M TEAA (pH 7.0)-50% acetonitrile; linear gradient from 20 to 25% B in 20 min at ambient temperature; flow-rate, 1 ml/min.

Fig. 4. Reversed-phase-ion-pair HPLC of oligo-(dA)s. Column: LiChrocart cartridge (25 cm × 4.6 mm I.D.) filled with LiChrosorb RP-8. Eluents: A, 50 mM potassium phosphate (pH 5.9)-2 mM TBAHSO4; B, 50 mM potassium phosphate (pH 6.5)-2 mM TBAHSO₄-60% acetonitrile; linear gradient from 15 to 80% B in 60 min at ambient temperature; flow-rate, 1.0 ml/min.

reversed-phase-ion-pair chromatogram of a mixture of d(GGAATTCC) and d(CGAATTCG). Although the retention time is longer than in the reversed-phase chromatography shown in Fig. 3, the separation is not as good.

Figs. 7 and 8 show chromatograms of mixtures of oligo-(dA)s and oligo-(dT)s separated by strong anion-exchange HPLC. The retention times under the chosen conditions are nearly constant with respect to base composition, e.g., $p(dA)_{10}$ is eluted at the same position as $p(dT)_{10}$. The resolution of oligonucleotides of increasing chain length is excellent up to a chain length of about 20. The high resolution is also demonstrated in Fig. 9, which shows the chromatogram of the separation of commercial oligo- $(dA)_n$. More than 30 discrete peaks can be seen. The usefulness of ion-exchange HPLC for the separation of longer oligonucleotides is particularly apparent from Fig. 10. In the HPLC run shown the mixture of products resulting from the T4 DNA ligase catalysed oligomerization of pd(GGAATTCC) was separ-

> CUREVAC EX2029 Page 5

DOCKET



Explore Litigation Insights

Docket Alarm provides insights to develop a more informed litigation strategy and the peace of mind of knowing you're on top of things.

Real-Time Litigation Alerts



Keep your litigation team up-to-date with **real-time** alerts and advanced team management tools built for the enterprise, all while greatly reducing PACER spend.

Our comprehensive service means we can handle Federal, State, and Administrative courts across the country.

Advanced Docket Research



With over 230 million records, Docket Alarm's cloud-native docket research platform finds what other services can't. Coverage includes Federal, State, plus PTAB, TTAB, ITC and NLRB decisions, all in one place.

Identify arguments that have been successful in the past with full text, pinpoint searching. Link to case law cited within any court document via Fastcase.

Analytics At Your Fingertips



Learn what happened the last time a particular judge, opposing counsel or company faced cases similar to yours.

Advanced out-of-the-box PTAB and TTAB analytics are always at your fingertips.

API

Docket Alarm offers a powerful API (application programming interface) to developers that want to integrate case filings into their apps.

LAW FIRMS

Build custom dashboards for your attorneys and clients with live data direct from the court.

Automate many repetitive legal tasks like conflict checks, document management, and marketing.

FINANCIAL INSTITUTIONS

Litigation and bankruptcy checks for companies and debtors.

E-DISCOVERY AND LEGAL VENDORS

Sync your system to PACER to automate legal marketing.

