

# Haptic feedback in robot-assisted minimally invasive surgery

Allison M. Okamura

Johns Hopkins University, Department of Mechanical Engineering, Laboratory for Computational Sensing and Robotics, Baltimore, Maryland, USA

Correspondence to Allison M. Okamura, Department of Mechanical Engineering, Laboratory for Computational Sensing and Robotics, 3400 N. Charles St., Baltimore, MD 21218, USA  
Tel: +1 410 516 7266; fax: +1 410 516 7254;  
e-mail: aokamura@jhu.edu

**Current Opinion in Urology** 2009, 19:102–107

## Purpose of review

Robot-assisted minimally invasive surgery (RMIS) holds great promise for improving the accuracy and dexterity of a surgeon and minimizing trauma to the patient. However, widespread clinical success with RMIS has been marginal. It is hypothesized that the lack of haptic (force and tactile) feedback presented to the surgeon is a limiting factor. This review explains the technical challenges of creating haptic feedback for robot-assisted surgery and provides recent results that evaluate the effectiveness of haptic feedback in mock surgical tasks.

## Recent findings

Haptic feedback systems for RMIS are still under development and evaluation. Most provide only force feedback, with limited fidelity. The major challenge at this time is sensing forces applied to the patient. A few tactile feedback systems for RMIS have been created, but their practicality for clinical implementation needs to be shown. It is particularly difficult to sense and display spatially distributed tactile information. The cost–benefit ratio for haptic feedback in RMIS has not been established.

## Summary

The designs of existing commercial RMIS systems are not conducive for force feedback, and creative solutions are needed to create compelling tactile feedback systems. Surgeons, engineers, and neuroscientists should work together to develop effective solutions for haptic feedback in RMIS.

## Keywords

force, haptics, minimally invasive surgery, robotics, tactile

Curr Opin Urol 19:102–107  
© 2009 Wolters Kluwer Health | Lippincott Williams & Wilkins  
0963-0643

## Introduction

Haptics generally describes touch feedback, which may include kinesthetic (force) and cutaneous (tactile) feedback. In manual minimally invasive surgery (MIS), surgeons feel the interaction of the instrument with the patient via a long shaft, which eliminates tactile cues and masks force cues. Some studies have linked the lack of significant haptic feedback in MIS to increased intraoperative injury [1]. In teleoperated robot-assisted minimally invasive surgery (RMIS), all natural haptic feedback is eliminated because the surgeon no longer manipulates the instrument directly. The lack of effective haptic feedback is often reported by surgeons and robotics researchers alike to be a major limitation of current RMIS systems.

## Haptic technology

In the robotics and virtual reality literature, haptics is broadly defined as real and simulated touch interactions between robots, humans, and real, remote, or simulated environments, in various combinations. The goal of haptic technology in robot-assisted minimally invasive surgery is to provide ‘transparency’, in which the surgeon

does not feel as if he is operating a remote mechanism, but rather that his own hands are contacting the patient. This requires artificial haptic sensors on the patient-side robot to acquire haptic information, and haptic displays to convey the information to the surgeon (Fig. 1). We categorize haptics as kinesthetic (related to forces and positions of the muscles and joints) and/or cutaneous (tactile; related to the skin) in nature. Haptics includes force, distributed pressure, temperature, vibrations, and texture, which are in some cases difficult to model and quantify, let alone acquire and display. To provide myriad haptic information to the surgeon without sacrificing the maneuverability and dexterity afforded by the RMIS system is a major technical challenge. Simultaneously, the robot components, particularly disposable instruments, must remain low cost and robust.

As a technical field, haptics research has been active for several decades. In the 1990s, haptics research expanded significantly with the availability of high-fidelity, commercially available force feedback systems from companies such as SensAble Technologies, Inc. (Woburn, Massachusetts, USA) and Immersion, Inc. (San Jose, California, USA). Currently, much of the force feedback research focuses on developing practical


0963-0643 © 2009 Wolters Kluwer Health | Lippincott Williams & Wilkins

DOI:10.1097/MOU.0b013e32831a478c

Copyright © Lippincott Williams & Wilkins. Unauthorized reproduction of this article is prohibited.

**Ethicon Exhibit 2018.001**  
**Intuitive v. Ethicon**

**Figure 1** The main components of a teleoperated robot for minimally invasive surgery with multimodal haptic feedback



Both force and tactile feedback are included in the model, and graphical display (one method of sensory substitution) is shown as a possible alternative to direct haptic feedback.


systems for application in fields such as entertainment, education, training, medicine and dentistry, and rehabilitation. Although researchers have studied tactile feedback for many years, there is currently no commercially available tactile display system that provides distributed information to the skin in a compact package feasible for applications. One aspect of tactile feedback that has proven effective in applications such as video games is vibration feedback, in which a single vibrating actuator is typically used to provide information about events such as making and breaking contact. Further reading about haptic technology includes books [2,3,4], tutorials [5,6,7], and research reviews [8–10]. Recent reviews of haptics in surgery are [11,12].

surgical instrument, and provide resolved forces to the hand via a force feedback device. Commercially available force sensors are very effective for measuring forces and torques in many teleoperation applications, but the surgical environment places severe constraints on size, geometry, cost, biocompatibility, and sterilizability. Although it is difficult to add force sensors to existing robotic instruments that were not designed with force sensing in mind, some researchers have had success on this front by creating specialized grippers that can attach to the jaws of existing instruments [13,14]. Another approach is to rethink the design of surgical instruments. The design of the force sensor can be integrated with that of the dexterous instrument [15,16,17], as shown in Fig. 2. For reasons of cost, biocompatibility, and sterilizability, the forces applied to the patient would ideally be estimated without using force sensors. For patient-side robots designed with low inertia and friction, the difference between the desired and actual position of the patient-side robot (where the desired

**Force feedback**

Kinesthetic or force feedback systems for RMIS typically measure or estimate the forces applied to the patient by the

**Figure 2** A robotic surgery system for two-hand manipulation with integrated force feedback and 3D vision, designed by researchers at DLR, Germany



The system consists of a specially designed dexterous force-sensing instrument, robotic arms and teleoperation controller, and haptic device commercially available from Force Dimension, Inc. (Lausanne, Switzerland) as the master manipulator. Original figures used with permission from B. Kuebler, DLR.

Copyright © Lippincott Williams & Wilkins. Unauthorized reproduction of this article is prohibited.

**Ethicon Exhibit 2018.002**  
**Intuitive v. Ethicon**

position is that of the master manipulator) is an indication for forces being applied to the environment. However, the fidelity of such systems are limited as there are dynamic forces present in most robots that are difficult to account for and often mask the relatively minute forces of interacting with the patient [18].

Direct force feedback to the surgeon's hands can use conventional force display technology, in which the motors of the master manipulator are programmed to recreate the forces sensed by the patient-side robot. A dexterous surgical robot typically has seven degrees of freedom of motion, including translational, rotational, and gripping. However, not all of those degrees of freedom are actuated on the master. That is, the system cannot provide force feedback in certain directions. The effects may be negligible or detrimental, depending on the directions of force feedback lost [19,20]. The dynamics of the master manipulator can also limit the accuracy of the force display, but a more fundamental limitation is the trade-off between system stability and transparency for force feedback. A perfectly transparent telemanipulator is not possible because it would require perfect models of the master and patient-side robot dynamics, zero time delays from computer processing and information transmission, and perfect environment force sensing or estimation. As one pushes toward the limit of transparency, small errors and delays in the system can cause uncontrollable oscillations in a 'closed-loop' teleoperator – this is known as instability and would be unacceptable in surgery. An alternative approach is to display force using sensory substitution to display force, including audio feedback [21], graphical feedback [22,23], or other forms of haptic feedback such as vibrotactile display [24]. Substantial information about environment properties and forces can be acquired by simply observing visually how the patient's tissue and materials such as suture respond to motions of the surgical instruments. A design guideline is that sensory substitution through graphical overlays should not distract from the surgeon's view of the patient via the endoscopic camera(s) [25\*\*].

In the last few years, several research groups have used the force sensing and feedback techniques described earlier to test the effectiveness of haptic feedback on surgeon performance and 'outcomes' in phantom patients. All the experiments to date are preclinical. (Current commercial systems that use haptic feedback include those of Hansen Medical and MAKO Surgical Corp; however, no data exists demonstrating the relative effectiveness of those systems with and without haptic feedback) Ortmaier, *et al.* [17\*] found that haptic feedback reduced unintentional injuries during a dissection task. However, operating time was longer than that of a manual intervention. Wagner and Howe [13\*\*] found that force feedback reduces potential tissue damage (as measured by the level of applied force)

for both surgeons and nonsurgeons, but only surgically trained individuals improve performance without a significant increase in trial time. They hypothesize that this is due to the interaction between visual-spatial motor abilities and the information contained in the mechanical interaction forces. Cao *et al.* [26\*] used a virtual environment to demonstrate, the surgeons performed a Transfer Place task faster and more accurately with haptics than without, even under cognitive load.

Mahvash *et al.* [27\*\*] used a modified da Vinci Surgical System to demonstrate that, in a palpation task, direct force feedback is superior to graphical force displays. Due to the limited fidelity of force feedback of their system (which did not use force sensors), users' identification of hard lumps in artificial tissue was only correct for models with significantly different mechanical properties between the lumps and surrounding tissue. Zhou *et al.* [28] did a study of MIS showing that, with trocar friction, one of the undesirable forces that arise in RMIS, surgeons' force perception was degraded and the time to detect contact was longer. When friction was present, experienced surgeons detected contact with tissue faster than novices. Compared to no force feedback, Reiley *et al.* [25\*\*] found that graphical displays of applied force during a knot-tying task reduced suture breakage and overall applied forces, and increased consistency of applied forces for inexperienced robot-assisted surgeons. In contrast to the direct force feedback results from [26\*], the results of Reiley *et al.* [25\*\*] suggest that graphical force feedback primarily benefits novices, with diminishing benefits among experienced surgeons. In a simple grasping task, Tholey *et al.* [29] found that providing both vision and force feedback leads to better tissue characterization than only vision or force feedback alone.

One would expect that better performance is achieved with direct force feedback than graphical feedback; sensory substitution systems are unnatural and thus have a longer learning curve, and direct force feedback provides physical constraints that helps a surgeon make the correct motions simply due to dynamic force balance [30\*\*]. There is an alternative to force feedback from the environment that provides such useful physical constraints: virtual fixtures. These are software-generated force and position signals applied to human operators in order to improve the safety, accuracy, and speed of robot-assisted manipulation tasks [31]. For example, a virtual 'wall' may be placed around a delicate anatomical structure to keep the surgical instruments from contacting it.

Although this article focuses on haptic feedback in actual surgeries, it is worth noting that the role of force feedback in training is a topic of current research. Some virtual reality simulators have proven effective in developing laparoscopic minimally invasive surgery (MIS) skills,

Copyright © Lippincott Williams & Wilkins. Unauthorized reproduction of this article is prohibited.

**Ethicon Exhibit 2018.003**  
**Intuitive v. Ethicon**

especially when force feedback is provided in early training [32,33]. However, accurate modeling of tissue–instrument interaction that gives rise to motions and forces relevant to outcomes is not yet achievable at computation rates that allow real-time interaction.

---

### Tactile feedback

Compared to force feedback, there has been relatively little work done in the area of tactile feedback for RMIS. In many surgical procedures, such as suture knot tying, force feedback is sufficient; the addition of contact location, finger pad deformation, and pressure distribution information may not be necessary [34]. However, palpation is one task for which deformation of the fingerpad seems to be particularly relevant [35,27<sup>\*\*</sup>], motivating the need for tactile feedback.

As in force feedback, tactile feedback systems require both a sensor and a display. The goal of tactile sensing in RMIS can be to detect local mechanical properties of tissue such as compliance, viscosity, and surface texture – all indications of the health of the tissue – or to obtain information that can be used directly for feedback to a human operator, such as pressure distribution or deformation over a contact area [36]. Constraints in sensor design include cost, size, geometry (for example, to fit within a laparoscopic grasper), biocompatibility, and surface finish to allow grasping. Many sensors require some deformation of the sensor in order to measure distributed information; this necessitates flexible coverings, which also remove detailed, local information. In addition, recording data from tactile sensors is difficult because they often include many individual sensing elements; wireless communications are possible, but power must still be cabled to the instrument tip. Examples of tactile sensors include arrays of capacitive sensors [37] and force-sensitive resistors [38], instrumented membranes [39], and micromachined piezoelectric arrays [40]. Companies that sell tactile array systems include Pressure Profile Systems, Inc. (Los Angeles, California, USA) and TekScan, Inc. (South Boston, Massachusetts, USA). Data relevant to tactile information can also be obtained through other means, such as laparoscopic ultrasound [41].

Tactile displays attempt to create the perception that the surgeon's fingertip is directly contacting the patient or surgical material such as suture. The most literal type of tactile display is an array of pins that are individually actuated (for example, [42]), so that their position commands are easily mapped from data from an array-type tactile sensor. Making such array-type displays for RMIS is very challenging due to size and weight constraints. The display must sit at the end of the master manipulator and not impede its maneuverability. Such pin displays developed for MIS and RMIS are actuated using shape-

memory alloys [43], micromotors [44], and pneumatic systems [45,46]. The latter method allows the most lightweight display to be attached to the master manipulator, but requires infrastructure for air pressure, which can be noisy, and has limited resolution. Little work has been done to combine kinesthetic and tactile information for surgery, but one study demonstrates that the ability to maintain an appropriate force in the remote environment is necessary for the surgeon to take full advantage of the spatially distributed force information from the tactile sensor [47]. Graphical displays of tactile data can also be very compelling, especially for diagnosis applications [48,49,50<sup>\*</sup>]. Most of the tactile sensors and displays developed have not been tested in RMIS systems. Due to the complexity of integrating distributed tactile information into RMIS, it may be useful in the future to consider clever 'tactile illusions' [51] and other display methods recently developed in the haptics research community.

---

### Conclusion

Haptic feedback for RMIS is currently being developed and evaluated in engineering laboratories, and further development is required before these techniques are ready for clinical testing. Because the fidelity of haptic feedback and surgical scenario (e.g., degrees of freedom and type of surgery) of each research system is different, the results regarding the effectiveness of haptic feedback in the literature are not completely consistent. Contributions by neuroscientists to our understanding of how humans perceive force and tactile information affects how we design haptic displays. Promising new developments in the haptic technology and neuroscience fields may yield more efficient, practical force and tactile displays in the future. To accomplish these goals, surgeons, engineers, and neuroscientists need to work together to develop and test effective haptic displays for RMIS.

---

### Acknowledgements

The author thanks current and former graduate students and postdoctoral fellows for their contributions to ideas presented in this review: J. Abbott, J. Gwilliam, K. Kuchenbecker, M. Mavash, C. Reiley, and L. Verner. This work was supported in part by National Institutes of Health grant R01 EB002004 and National Science Foundation grants IIS-0347464 and EEC-9731478.

---

### References and recommended reading

Papers of particular interest, published within the annual period of review, have been highlighted as:

- of special interest
- of outstanding interest

Additional references related to this topic can also be found in the Current World Literature section in this issue (p. 123).

- 1 Xin H, Zelek JS, Carnahan H. Laparoscopic surgery, perceptual limitations and force: a review. In: First Canadian Student Conference on Biomedical Computing, Kingston, Ontario, Canada; 2006. No. 144.
- 2 Burdea CG. Force and Touch Feedback for Virtual Reality. New York: Wiley Interscience; 1996.

Copyright © Lippincott Williams & Wilkins. Unauthorized reproduction of this article is prohibited.

**Ethicon Exhibit 2018.004**  
**Intuitive v. Ethicon**

- 3 Jones L, Lederman S. *Human Hand Function*. New York: Oxford University Press; 2006.
- 4 Lin MC, Otaduy MA, editors. *Haptic Rendering: Foundations, Algorithms, and Applications*. London: AK Peters, Ltd; 2008. Provides a current and in-depth review of the field of haptics, focusing on virtual environments. Medical applications are covered.
- 5 Hayward V, MacLean KE. Do it yourself haptics, Part-I. *IEEE Robot. Autom Mag* 2007; 14:88–104.
- 6 MacLean KE, Hayward V. Do It Yourself Haptics, Part-II. *IEEE Robot Autom Mag* 2008; 15:104–119.
- 7 Hannaford B, Okamura AM. Chapter 30: Haptics. In: Siciliano B, Khatib O, editors. *Handbook of Robotics*. New York: Springer; 2008. Provides a brief overview of the field of haptics suitable for a scientifically literate audience.
- 8 Hayward V, Astley OR, Cruz-Hernandez M, et al. Haptic interfaces and devices. *Sensor Rev* 2004; 24:16–29.
- 9 Salisbury K, Conti F, Barbagli F. Haptic rendering: introductory concepts. *IEEE Comput Graphics Appl* 2004; 24:24–32.
- 10 Robles-De-La-Torre G. The importance of the sense of touch in virtual and real environments. *IEEE Multimedia* 2006; 13:24–30.
- 11 Westebring - van der Putten EP, Goossens RHM, Jakimowicz JJ, Dankelman J. Haptics in minimally invasive surgery: a review. *Minim Invasive Ther Allied Technol* 2008; 17:3–16. Reviews the literature from 1985 to 2007 for haptics in both MIS and RMIS.
- 12 Puanngali P, Althoefer K, Seneviratne LD, et al. State-of-the-art in force and tactile sensing for minimally invasive surgery. *IEEE Sensors Journal* 2008; 8:371–381. Reviews recent developments of haptic sensing in MIS.
- 13 Wagner CR, Howe RD. Force Feedback Benefit Depends on Experience in Multiple Degree of Freedom Robotic Surgery Task. *IEEE Transactions on Robotics* 2007; 23:1235–1240. Demonstrates a complete teleoperation system for force feedback RMIS and shows that, while all users improve accuracy with force feedback, only experienced surgeons do not take longer to perform a surgical task.
- 14 Dargahi J, Sedaghati R, Singh H, Najarian S. Modeling and testing of an endoscopic piezoelectric-based tactile sensor. *Mechatronics* 2007; 17: 462–467.
- 15 Kuebler B, Seibold U, Hirzinger G. Development of actuated and sensor integrated forceps for minimally invasive robotic surgery. *Int J Med Robotics Comput Assist Surg* 2005; 1:96–107.
- 16 Zemiti N, Morel G, Ortmaier T, Bonnet N. Mechatronic design of a new robot for force control in minimally invasive surgery. *IEEE/ASME Transactions on Mechatronics* 2007; 12:143–153.
- 17 Ortmaier T, Deml B, Kuebler B, et al. Robot assisted force feedback surgery. In: Ferre M, Buss M, Aracil R, et al., editors. *Advances in Telerobotics*, Springer Tracts in Advanced Robotics (STAR) Vol. 31. New York: Springer; 2007. pp. 341–358. With a custom teleoperation system designed for RMIS, the authors showed that haptic feedback reduced unintentional injuries during a dissection task. However, operating time was longer than that of a manual intervention.
- 18 Mahvash M, Okamura AM. Friction compensation for enhancing transparency of a teleoperator with compliant transmission. *IEEE Transactions on Robotics* 2007; 23:1240–1246.
- 19 Semere W, Kitagawa M, Okamura AM. Teleoperation with sensor/actuator asymmetry: task performance with partial force feedback. In: 12th Symposium on Haptic Interfaces for Virtual Environments and Teleoperator Systems 2004; pp. 121–127.
- 20 Verner LN, Okamura AM. Effects of translational and gripping force feedback are decoupled in a 4-degree-of-freedom telemanipulator. In: Second Joint Eurohaptics Conference and Symposium on Haptic Interfaces for Virtual Environment and Teleoperator Systems (World Haptics); 2007, pages 286–291.
- 21 Kitagawa M, Dokko D, Okamura AM, Yuh DD. Effect of Sensory Substitution on Suture Manipulation Forces for Robotic Surgical Systems. *J Thorac Cardiovasc Surg* 2005; 129:151–158.
- 22 Okamura AM. Methods for haptic feedback in teleoperated robot-assisted surgery. *Industrial Robot* 2004; 31:499–508.
- 23 Tavakoli M, Aziminejad A, Patel RV, Moallem M. Methods and Mechanisms for Contact Feedback in a Robot-Assisted Minimally Invasive Environment. *Surg Endosc Other Interv Tech* 2006; 10:1570–1579.
- 24 Schoonmaker RE, Cao CGL. Vibrotactile force feedback system for minimally invasive surgical procedures. In: IEEE Conference on Systems, Man, and Cybernetics 2006; pp. 2464–2469.
- 25 Reiley CE, Akinbiyi T, Burschka D, et al. Effects of visual force feedback on robot-assisted surgical task performance. *J Thorac Cardiovasc Surg* 2008; 135:196–202. Authors showed that graphical displays of applied force during a knot-tying task reduced suture breakage and overall applied forces, and increased consistency of applied forces for inexperienced robot-assisted surgeons.
- 26 Cao CGL, Zhou M, Jones DB, Schweitzberg SD. Can surgeons think and operate with haptics at the same time? *J Gastrointest Surg* 2007; 11:1564–1569. Authors use a virtual environment to show that force feedback increases accuracy.
- 27 Mahvash M, Gwilliam J, Agarwal R, et al. Force-feedback surgical teleoperator: controller design and palpation experiments. In: 16th Symposium on Haptic Interfaces for Virtual Environments and Teleoperator Systems 2008; pp. 465–471. Experimental results with a custom force feedback teleoperator based on da Vinci Surgical System hardware showed that direct force feedback is better for palpation than graphical force feedback, but that attempting force feedback without using force sensors limits the system fidelity.
- 28 Zhou M, Perreault J, Schweitzberg SD, Cao CGL. Effects of experience on force perception threshold in minimally invasive surgery. *Surg Endosc* 2008; 22:510–515.
- 29 Tholey G, Desai JP, Castellanos AE. Force feedback plays a significant role in minimally invasive surgery: results and analysis. *Annals of Surgery* 2005; 241:102–109.
- 30 Wagner CR, Stylopoulos N, Jackson PG, Howe RD. The benefit of force feedback in surgery: examination of blunt dissection. Presence: teleoperators and virtual environments 2007; 16:252–262. Authors show that increased force feedback results in improved accuracy for a mock blunt dissection task. However, haptic feedback lengthens operation time for novice surgeons.
- 31 Abbott JJ, Marayong P, Okamura AM. Haptic Virtual Fixtures for Robot-Assisted Manipulation. In: Thrun S, Durrant-Whyte H, Brooks R, editors. *Robotics Research*, Springer Tracts in Advanced Robotics, Vol. 28. 2007; pp. 49–64.
- 32 Basdogan C, De S, Kim J, Muniyandi M. Haptics in minimally invasive surgical simulation and training. *IEEE Computer Graphics Applications* 2004; 24:56–64.
- 33 Ström P, Hedman L, Särnå L, et al. Early exposure to haptic feedback enhances performance in surgical simulator training: a prospective randomized crossover study in surgical residents. *Surg Endosc* 2006; 20:1383–1388.
- 34 Kitagawa M, Okamura AM, Bethea BT, et al. Analysis of suture manipulation forces for teleoperation with force feedback. In: Proceedings of the Fifth International Conference on Medical Image Computing and Computer Assisted Intervention: MICCAI 2002, Lecture Notes in Computer Science, Vol. 2488. Dohi T, Kikinis R, editors. 2002; pp. 155–162.
- 35 Peine WJ, Howe RD. Do humans sense finger deformation or distributed pressure to detect lumps in soft tissue? In: Proc. of the ASME Dynamic Systems and Control Division, ASME International Mechanical Engineering Congress and Exposition, DSC-Vol. 64. 1998; pp. 273–278.
- 36 Eltaib MEH, Hewit JR. Tactile sensing technology for minimal access surgery: a review. *Mechatronics* 2003; 13:1163–1177.
- 37 Pawluc DTV, Son JS, Wellman PS, et al. A Distributed Pressure Sensor for Biomechanical Measurements. *ASME J Biomech Eng* 1998; 102:302–305.
- 38 Schostek S, Ho CN, Kalanovic D, Schurr MO. Artificial tactile sensing in minimally invasive surgery: a new technical approach. *Minim Invasive Ther Allied Technol* 2006; 15:296–304.
- 39 Dargahi J, Najarian S. A supported membrane type sensor for medical tactile mapping. *Sensor Rev* 2004; 24:284–297.
- 40 Dargahi J, Parameswaran M, Payandeh S. A micromachined piezoelectric tactile sensor for an endoscopic grasper – theory, fabrication and experiments. *J Microelectromechanical Syst* 2000; 9:329–335.
- 41 Leven J, Burschka D, Kumar R, et al. DaVinci Canvas: a telerobotic surgical system with integrated, robot-assisted, laparoscopic ultrasound capability. In: Proceedings of the Eighth International Conference on Medical Image Computing and Computer Assisted Intervention: MICCAI, Lecture Notes in Computer Science (Vol. 3750) 2005; pp. 811–818.
- 42 Killebrew JH, Bensmaia SJ, Dammann JF, et al. A dense array stimulator to generate arbitrary spatio-temporal tactile stimuli. *J Neurosci Methods* 2007; 161:62–74.
- 43 Howe RD, Peine WJ, Kontarinis DA, Son JS. Remote palpation technology. *IEEE Eng Med Biol* 1995; 14:318–323.
- 44 Ottermo MV, Stavadahl O, Johansen TA. Electromechanical design of a miniature tactile shape display for minimally invasive surgery. In: First Joint Eurohaptics Conference and Symposium on Haptic Interfaces for Virtual Environment and Teleoperator Systems (World Haptics) 2005; pp. 561–562.

Copyright © Lippincott Williams & Wilkins. Unauthorized reproduction of this article is prohibited.

**Ethicon Exhibit 2018.005**  
**Intuitive v. Ethicon**

# Explore Litigation Insights

Docket Alarm provides insights to develop a more informed litigation strategy and the peace of mind of knowing you're on top of things.

## Real-Time Litigation Alerts



Keep your litigation team up-to-date with **real-time alerts** and advanced team management tools built for the enterprise, all while greatly reducing PACER spend.

Our comprehensive service means we can handle Federal, State, and Administrative courts across the country.

## Advanced Docket Research



With over 230 million records, Docket Alarm's cloud-native docket research platform finds what other services can't. Coverage includes Federal, State, plus PTAB, TTAB, ITC and NLRB decisions, all in one place.

Identify arguments that have been successful in the past with full text, pinpoint searching. Link to case law cited within any court document via Fastcase.

## Analytics At Your Fingertips



Learn what happened the last time a particular judge, opposing counsel or company faced cases similar to yours.

Advanced out-of-the-box PTAB and TTAB analytics are always at your fingertips.

## API

Docket Alarm offers a powerful API (application programming interface) to developers that want to integrate case filings into their apps.

## LAW FIRMS

Build custom dashboards for your attorneys and clients with live data direct from the court.

Automate many repetitive legal tasks like conflict checks, document management, and marketing.

## FINANCIAL INSTITUTIONS

Litigation and bankruptcy checks for companies and debtors.

## E-DISCOVERY AND LEGAL VENDORS

Sync your system to PACER to automate legal marketing.