

UNITED STATES PATENT AND TRADEMARK OFFICE

BEFORE THE PATENT TRIAL AND APPEAL BOARD

MEDTRONIC, INC., AND MEDTRONIC VASCULAR, INC.,
Petitioner,

v.

TELEFLEX INNOVATIONS S.À.R.L.,
Patent Owner.

IPR2020-00133
Patent RE45,760 E

Before SHERIDAN K. SNEDDEN, JON B. TORNQUIST, and
CHRISTOPHER G. PAULRAJ, *Administrative Patent Judges*.

SNEDDEN, *Administrative Patent Judge*.

DECISION
Denying Institution of *Inter Partes* Review
35 U.S.C. § 314

I. INTRODUCTION

Medtronic, Inc., and Medtronic Vascular, Inc. (“Petitioner”) filed a Petition for *inter partes* review of claims 25–42, 44, and 47 of U.S. Patent No. RE45,760 E (“the ’760 patent,” Ex. 1201). Paper 1 (“Pet.”). Teleflex Medical Devices S.A.R.L. (“Patent Owner”) filed a Preliminary Response. Paper 10 (“Prelim. Resp.”). Pursuant to our authorization, Petitioner filed a Reply addressing its burden on secondary considerations and reduction to practice, and Patent Owner filed a Sur-Reply addressing Petitioner’s burden on those issues. Paper 14; Paper 15. Also pursuant to our authorization, Petitioner filed another Reply (Paper 17), and Patent Owner filed another Sur-Reply (Paper 18) addressing the factors for discretionary denial under 35 U.S.C. § 314(a).

To institute an *inter partes* review, we must determine that the information presented in the Petition shows “a reasonable likelihood that the petitioner would prevail with respect to at least 1 of the claims challenged in the petition.” 35 U.S.C. § 314(a) (2012). After considering the evidence and arguments presented in the Petition and Preliminary Response, we determine that Petitioner has not satisfied its burden under § 314. Thus, we do not institute an *inter partes* review.

A. Real Parties in Interest

Petitioner identifies its real parties-in-interest as Medtronic, Inc. and Medtronic Vascular, Inc., and notes that “Medtronic plc is the ultimate parent of both entities.” Pet. 5. Patent Owner identifies its real parties-in-interest as Teleflex Medical Devices S.A.R.L.; Vascular Solutions LLC; Arrow International, Inc.; and Teleflex LLC. Paper 4, 2. Patent Owner also notes that “Teleflex Incorporated is the ultimate parent of the entities listed above.”

B. Related Matters

The '760 patent is also at issue in IPR2020-00132 and IPR2020-00134. Paper 4, 2–3; Pet. 5. We instituted *inter partes* review in IPR2020-00132 on June 8, 2020. IPR2020-00132, Paper 22.

The parties indicate that the '760 patent is the subject of litigation in *Vascular Solutions LLC, et al. v. Medtronic, Inc., et al.* No. 19-cv-01760 (D. Minn. filed July 2, 2019), and *QXMedical, LLC v. Vascular Solutions, LLC*, No. 17-cv-01969 (D. Minn., filed June 8, 2017). Pet. 5; Paper 4, 2.

The '760 patent is a reissue of U.S. Pat. No. 8,292,850 (“the '850 patent”). The '850 patent was the subject of two previous *inter partes* reviews: IPR2014-00762, filed May 16, 2014, and terminated August 11, 2014 by way of joint motion to terminate, and IPR2014-00763, filed May 16, 2014, and terminated August 11, 2014, by way of joint motion to terminate. Pet. 5; Paper 4, 2–3. The '850 patent was also at issue in the U.S. District Court for the District of Minnesota in *Vascular Solutions, Inc. v. Boston Scientific Corp.*, No. 13-cv-01172 (D. Minn., filed May 16, 2013). *Id.*

C. The '760 Patent

1. Specification

The subject matter claimed in the '760 patent is directed to a device for use with a standard guide catheter. Ex. 1201, 13:36–17:13. Figures 1 and 5 of the '760 patent, reproduced below, depict a coaxial guide catheter and a tapered inner catheter.




Figure 1 of the '760 patent




Figure 5 of the '760 patent

As shown in Figures 1 and 5, above, coaxial guide catheter assembly 10 includes coaxial guide catheter 12 and tapered inner catheter 14. *Id.* at 6:37–39. Coaxial guide catheter 12 generally includes tip portion 16, reinforced portion 18, and rigid portion 20. *Id.* at 6:40–41. Tip portion 16 generally includes bump tip 22 and marker band 24. *Id.* at 6:44–45. Bump tip 22 includes taper 26 and is relatively flexible. *Id.* at 6:45–46. Marker band 24 is formed of a radiopaque material such as platinum/iridium alloy. *Id.* at 6:49–50. Tapered inner catheter tip 42 includes tapered portion 46 at a distal end thereof, and straight portion 48. *Id.* at 7:22–23. Both tapered portion 46 and straight portion 48 are pierced by lumen 50. *Id.* at 7:23–24. Tapered inner catheter 14 may also include clip 54 at a proximal end thereof to releasably join tapered inner catheter 14 to coaxial guide catheter 12. *Id.* at 7:27–29. Thus, tapered inner catheter 14 is keyed to coaxial guide catheter 12. *Id.* at 7:29–30.

2. *Illustrative Claim*

Independent claim 25, reproduced below, is illustrative of the challenged claims.

25. A system, comprising:

a guide catheter configured to be advanceable through a main blood vessel to a position adjacent an ostium of a coronary artery, the guide catheter having a lumen extending from a hemostatic valve at a proximal end of the guide catheter to a distal end of the guide catheter that is adapted to be positioned adjacent the ostium of the coronary artery; and

a guide extension catheter configured to be partially advanceable through the guide catheter and into the coronary artery, the guide extension catheter having a length such that a distal end of the guide extension catheter is extendable through the lumen and beyond the distal end of the guide catheter, and a proximal end of the guide extension catheter is extendable through the hemostatic valve at the proximal end of the guide catheter,

the guide extension catheter including, in a proximal to distal direction, a substantially rigid segment, a segment defining a side opening, and *a tubular structure defining a lumen coaxial and in fluid communication with the lumen of the guide catheter*, the lumen of the tubular structure having a length that is shorter than the length of the lumen of the guide catheter and having a uniform cross-sectional inner diameter that is not more than one French size smaller than the cross-sectional inner diameter of the lumen of the guide catheter, the side opening extending for a distance along a longitudinal axis of the segment defining the side opening and accessible from a longitudinal side defined transverse to the longitudinal axis, and the side opening and the lumen of the tubular structure configured to receive one or more stents or balloon catheters when the segment defining the side opening and a proximal end portion of the tubular structure are positioned within the lumen of the guide catheter and the distal end of the guide extension catheter extends beyond the distal end of the guide catheter;

Explore Litigation Insights

Docket Alarm provides insights to develop a more informed litigation strategy and the peace of mind of knowing you're on top of things.

Real-Time Litigation Alerts



Keep your litigation team up-to-date with **real-time alerts** and advanced team management tools built for the enterprise, all while greatly reducing PACER spend.

Our comprehensive service means we can handle Federal, State, and Administrative courts across the country.

Advanced Docket Research



With over 230 million records, Docket Alarm's cloud-native docket research platform finds what other services can't. Coverage includes Federal, State, plus PTAB, TTAB, ITC and NLRB decisions, all in one place.

Identify arguments that have been successful in the past with full text, pinpoint searching. Link to case law cited within any court document via Fastcase.

Analytics At Your Fingertips



Learn what happened the last time a particular judge, opposing counsel or company faced cases similar to yours.

Advanced out-of-the-box PTAB and TTAB analytics are always at your fingertips.

API

Docket Alarm offers a powerful API (application programming interface) to developers that want to integrate case filings into their apps.

LAW FIRMS

Build custom dashboards for your attorneys and clients with live data direct from the court.

Automate many repetitive legal tasks like conflict checks, document management, and marketing.

FINANCIAL INSTITUTIONS

Litigation and bankruptcy checks for companies and debtors.

E-DISCOVERY AND LEGAL VENDORS

Sync your system to PACER to automate legal marketing.