

# **Pulse Oximeters - Premarket Notification Submissions [510(k)s]**

---

## **Guidance for Industry and Food and Drug Administration Staff**

**Document issued on: March 4, 2013**

**This document supersedes Non-invasive Pulse Oximeter General Guidance  
Document, September 7, 1992**

**The draft of this document was issued on July 19, 2007.**

For questions regarding this document contact Neel Patel at 301-796-5580 or  
neel.patel@fda.hhs.gov.



**U.S. Department of Health and Human Services  
Food and Drug Administration  
Center for Devices and Radiological Health  
Office of Device Evaluation  
Division of Anesthesiology, General Hospital,  
Infection Control, and Dental Devices  
Anesthesiology and Respiratory Devices Branch**

## Preface

### Public Comment

You may submit written comments and suggestions at any time for Agency consideration to the Division of Dockets Management, Food and Drug Administration, 5630 Fishers Lane, rm. 1061, (HFA-305), Rockville, MD, 20852. Submit electronic comments to <http://www.regulations.gov>. Identify all comments with the docket number listed in the notice of availability that publishes in the *Federal Register*. Comments may not be acted upon by the Agency until the document is next revised or updated.

### Additional Copies

Additional copies are available from the Internet. You may also send an e-mail request to [dsmica@fda.hhs.gov](mailto:dsmica@fda.hhs.gov) to receive an electronic copy of the guidance or send a fax request to 301-827-8149 to receive a hard copy. Please use the document number (1605) to identify the guidance you are requesting.

## Table of Contents

<b>1. INTRODUCTION.....</b>	<b>1</b>
<b>2. SCOPE .....</b>	<b>1</b>
<b>3. DEVICE DESCRIPTION .....</b>	<b>2</b>
<b>4. DEVICE PERFORMANCE.....</b>	<b>4</b>
4.1 ACCURACY OF PULSE OXIMETERS.....	4
4.2 ALARMS .....	8
4.3 DISPLAY VALUES, OUTPUTS, AND INDICATORS.....	8
4.4 SATURATION PULSE INFORMATION SIGNAL .....	9
<b>5. SOFTWARE INFORMATION .....</b>	<b>9</b>
<b>6. ELECTRICAL, MECHANICAL, AND ENVIRONMENTAL SAFETY .....</b>	<b>9</b>
<b>7. ELECTROMAGNETIC COMPATIBILITY .....</b>	<b>9</b>
<b>8. BIOCOMPATIBILITY .....</b>	<b>9</b>
<b>9. CLEANING, DISINFECTION, AND STERILIZATION .....</b>	<b>10</b>
9.1 REUSE INSTRUCTIONS AND VALIDATION.....	10
9.2 STERILIZATION DOCUMENTATION .....	10
<b>10. LABELING .....</b>	<b>10</b>
10.1 INTENDED USE .....	11
10.2 INSTRUCTIONS FOR USE .....	11
10.3 DEVICE SPECIFICATIONS.....	11
10.4 PACKAGE LABELING .....	12
<b>11. SUBMISSIONS FOR REPROCESSED SINGLE-USE SENSORS.....</b>	<b>12</b>
11.1 IDENTIFICATION OF COMPONENTS AND USES .....	13
11.2 PERFORMANCE TESTING .....	13
11.3 CLEANING METHODS AND VALIDATION INFORMATION.....	13
11.4 DISINFECTION AND/OR STERILIZATION VALIDATION INFORMATION .....	14
11.5 DEVICES NOT PROVIDED STERILE.....	14
11.6 DEVICES PROVIDED STERILE .....	14
<b>APPENDIX – LIST OF REFERENCES .....</b>	<b>15</b>

# Pulse Oximeters - Premarket Notification Submissions [510(k)s]

---

## Guidance for Industry and Food and Drug Administration Staff

*This guidance represents the Food and Drug Administration's (FDA's) current thinking on this topic. It does not create or confer any rights for or on any person and does not operate to bind FDA or the public. You can use an alternative approach if the approach satisfies the requirements of the applicable statutes and regulations. If you want to discuss an alternative approach, contact the FDA staff responsible for implementing this guidance. If you cannot identify the appropriate FDA staff, call the appropriate number listed on the title page of this guidance.*

### 1. Introduction

FDA has developed this guidance document to assist industry in preparing premarket notifications (510(k)s) for pulse oximeters. These devices are intended for non-invasive measurement of the arterial blood oxygen saturation and pulse rate.

FDA's guidance documents, including this guidance, do not establish legally enforceable responsibilities. Instead, guidances describe the Agency's current thinking on a topic and should be viewed only as recommendations, unless specific regulatory or statutory requirements are cited. The use of the word *should* in Agency guidances means that something is suggested or recommended, but not required.

### 2. Scope

The scope of this document is limited to the Class II devices, Oximeter and Ear oximeter, classified under the following regulations:

**21 CFR 870.2700 – Oximeter** (product codes: DQA (Oximeter) and NLF (Oximeter, Reprocessed))

An oximeter is a device used to transmit radiation at a known wavelength(s) through blood and to measure the blood oxygen saturation based on the amount of reflected or scattered radiation. It may be used alone or in conjunction with a fiberoptic oximeter catheter.

## *Contains Nonbinding Recommendations*

This guidance does not address oximeters in product codes MUD (tissue saturation oximeter), NMD (reprocessed tissue saturation oximeter), or MMA (fetal pulse oximeter).

### **21 CFR 870.2710 – Ear Oximeter** (product code: DPZ (Ear oximeter))

An ear oximeter is an extravascular device used to transmit light at a known wavelength(s) through blood in the ear. The amount of reflected or scattered light as indicated by this device is used to measure the blood oxygen saturation.

These classification regulations group together all oximeters intended to measure blood oxygen saturation. The regulations at 21 CFR 870.2700 and 870.2710 include devices using reflectance, transmittance, and fiber optic technologies, which are collectively referred to as pulse oximeters for the purpose of this guidance. The terms “transmittance” and “reflectance” refer to the sensor geometry and are not related to the principles of pulse oximetry and how the light is absorbed by hemoglobin.

This guidance document pertains to non-invasive pulse oximeters to measure arterial blood oxygen saturation (SpO<sub>2</sub>) and pulse rate based on the amount of transmitted, reflected and scattered light through various application sites (including, but not limited to finger, ear, foot, hand, forehead, back, and nose). These devices are limited to prescription use. These pulse oximeters may be continuous or spot-checking devices and either stand-alone or multi-parameter modules. These devices are typically labeled with a general indication for non-invasive measurement of blood oxygen saturation. A manufacturer that wishes to seek a specific clinical indication for use of a pulse oximeter, for example to screen for or diagnose a disease or condition, should submit clinical safety and effectiveness data to support the specific indication. A clinical evaluation of a new intended use of a legally marketed device may require an [Investigational Device Exemption \(IDE\)](http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/InvestigationalDeviceExemptionIDE/default.htm) under 21 CFR Part 812 (<http://www.fda.gov/MedicalDevices/DeviceRegulationandGuidance/HowtoMarketYourDevice/InvestigationalDeviceExemptionIDE/default.htm>) before the clinical study is initiated. Should an IDE be necessary, FDA suggests that manufacturers take advantage of the Pre-Submission Program (see “[Draft Guidance for Industry and FDA Staff Medical Devices: The Pre-Submission Program and Meetings with FDA Staff](#),” issued July 13, 2012,<sup>1</sup> which when final will represent FDA’s current thinking on this topic) to obtain input from FDA on their study plan.

## **3. Device Description**

We recommend you identify your device by the applicable regulation number and product code indicated in Section 2 above and include the information described below.

### **Intended Use**

We recommend you clarify if the device [and accessories] is intended:

---

<sup>1</sup> Web addresses for all guidance documents referenced within can be found in the “List of References” at the end of this document

# Explore Litigation Insights

Docket Alarm provides insights to develop a more informed litigation strategy and the peace of mind of knowing you're on top of things.

## Real-Time Litigation Alerts



Keep your litigation team up-to-date with **real-time alerts** and advanced team management tools built for the enterprise, all while greatly reducing PACER spend.

Our comprehensive service means we can handle Federal, State, and Administrative courts across the country.

## Advanced Docket Research



With over 230 million records, Docket Alarm's cloud-native docket research platform finds what other services can't. Coverage includes Federal, State, plus PTAB, TTAB, ITC and NLRB decisions, all in one place.

Identify arguments that have been successful in the past with full text, pinpoint searching. Link to case law cited within any court document via Fastcase.

## Analytics At Your Fingertips



Learn what happened the last time a particular judge, opposing counsel or company faced cases similar to yours.

Advanced out-of-the-box PTAB and TTAB analytics are always at your fingertips.

## API

Docket Alarm offers a powerful API (application programming interface) to developers that want to integrate case filings into their apps.

## LAW FIRMS

Build custom dashboards for your attorneys and clients with live data direct from the court.

Automate many repetitive legal tasks like conflict checks, document management, and marketing.

## FINANCIAL INSTITUTIONS

Litigation and bankruptcy checks for companies and debtors.

## E-DISCOVERY AND LEGAL VENDORS

Sync your system to PACER to automate legal marketing.