

Detection and quantification of mutations in the plasma of patients with colorectal tumors

Frank Diehl*, Meng Li*, Devin Dressman*, Yiping He*, Dong Shen*, Steve Szabo*, Luis A. Diaz, Jr., Steven N. Goodman*, Kerstin A. David†, Hartmut Juhl†, Kenneth W. Kinzler*, and Bert Vogelstein**

*Howard Hughes Medical Institute and The Sidney Kimmel Comprehensive Cancer Center, The Johns Hopkins Medical Institutions, 1650 Orleans Street, Baltimore, MD 21231; and †Indivumed, Center for Cancer Research at Israelitic Hospital, Orchideenstieg 14, 22297 Hamburg, Germany

Contributed by Bert Vogelstein, September 16, 2005

The early detection of cancers through analysis of circulating DNA could have a substantial impact on morbidity and mortality. To achieve this goal, it is essential to determine the number of mutant molecules present in the circulation of cancer patients and to develop methods that are sufficiently sensitive to detect these mutations. Using a modified version of a recently developed assay for this purpose, we found that patients with advanced colorectal cancers consistently contained mutant adenomatous polyposis coli (APC) DNA molecules in their plasma. The median number of APC DNA fragments in such patients was 47,800 per ml of plasma, of which 8% were mutant. Mutant APC molecules were also detected in >60% of patients with early, presumably curable colorectal cancers, at levels ranging from 0.01% to 1.7% of the total APC molecules. These results have implications for the mechanisms through which tumor DNA is released into the circulation and for diagnostic tests based on this phenomenon.

colorectal cancer | plasma DNA | tumor suppressor gene | circulating DNA | diagnosis

The probability of curing cancers through surgery alone is high in individuals whose primary tumors are detected at a relatively early stage. Such early detection is therefore one of the most promising approaches for limiting cancer morbidity and mortality in the future (1). At present, Pap smears can be used to detect cervical cancers, mammography can detect breast cancers, serum PSA (prostate-specific antigen) levels can signify the presence of prostate cancer, and colonoscopy and fecal occult blood tests can detect colon cancers (2). However, problems with sensitivity, specificity, cost, or compliance have complicated widespread implementation of many of these tests (3–5). Moreover, methods for the early detection of most other cancer types are not yet available.

The discovery of the genetic bases of neoplasia has led to new approaches to detect tumors noninvasively (6–8). Several of these approaches rely on the *ex vivo* detection of mutant forms of the oncogenes and tumor suppressor genes that are responsible for the initiation and progression of tumors. This approach was first used to detect bladder and colon tumors through examination of urine and stool, respectively (9, 10), and has since been used to detect several other tumor types (11–14). Because the mutant genes are not only “markers” for cancer but also the proximate causes of tumor growth (1), they have major conceptual advantages over conventional markers such as fecal occult blood or serum PSA. In particular, conventional markers are not pathogenically involved in the tumorigenic process and are much less specific for neoplasia than are mutations.

The evaluation of patient blood samples for mutant DNA molecules is a particularly attractive approach because such tests could detect many different forms of cancers. Additionally, blood can be easily obtained from patients during routine outpatient visits, and methods for preparing and storing plasma and serum are well known and reliable. Accordingly, numerous studies have attempted to identify abnormal forms or quantities of DNA in plasma or serum (6, 11–15). Unfortunately, the results of many of

these studies are contradictory. Some report high detection rates of cancers, and others report very low detection rates, despite the use of similar techniques and patient cohorts. Moreover, several studies have shown that loss of heterozygosity is routinely detectable in circulating DNA, even in patients with relatively nonaggressive tumors. To detect loss of heterozygosity in such samples, the neoplastic cells within a tumor must contribute >50% of the total circulating DNA.

The above studies, although promising, lead to several questions that must be answered to engender confidence in the use of circulating, abnormal DNA as a biomarker of malignancy. First, how many copies of a given gene fragment are present in the circulation in cancer patients? Second, what is the nature of this DNA (e.g., intact vs. degraded)? Third, what fraction of these gene fragments have an abnormal (e.g., mutant) DNA sequence? And, fourth, how does this fraction vary with stage of disease? To answer these questions, it was necessary to develop technologies that could simultaneously quantify the number of normal and mutant DNA molecules in a given sample, even when the fraction of mutant molecules was very small. In the current study, we employ such a technology to investigate circulating DNA in patients with colorectal tumors.

Materials and Methods

Sample Collection, DNA Extraction, and Sequencing. Detailed methods for these procedures are provided in the supporting information, which is published on the PNAS web site.

Real-Time PCR. Primers were designed to generate ≈100-bp amplicons that included one or more mutation sites. A universal tag (5'-TCCCGCGAAATTAATACGAC-3') was added to the 5' end of either the forward or reverse primer used to generate each amplicon. The sequences of these primers are listed in the supporting information. PCR was performed in 50- μ l reactions containing 10 μ l of 5 \times Phusion HF buffer, a 0.2 mM concentration of each dNTP, a 1 μ M concentration of each primer, 1:50,000 dilution of SYBR green I (Invitrogen), 1.5 units of Phusion DNA polymerase (NEB, Beverly, MA), and 15 μ l of purified plasma DNA (equivalent to 100 μ l of plasma) or genomic DNA purified from normal mononuclear cells of the blood of healthy volunteers. The amplifications were carried out with an iCycler (Bio-Rad) under the following conditions: 98°C for 1 min; 98°C for 10 s, 70°C for 10 s, and 72°C for 10 s 3 times; 98°C for 10 s, 67°C for 10 s, and 72°C for 10 s 3 times; 98°C for 10 s, 64°C for 10 s, and 72°C for 10 s 3 times;

Conflict of interest statement: Under a licensing agreement between EXACT Sciences and The Johns Hopkins University, K.W.K. and B.V. are entitled to a share of royalties received by the university on sales of products related to digital PCR. Under a licensing agreement between Agencourt Biosciences Corporation and The Johns Hopkins University, D.D., K.W.K., and B.V. are entitled to a share of royalties received by the university on sales of products related to the use of BEAMing for preparing templates for DNA sequencing. The terms of these arrangements are being managed by The Johns Hopkins University in accordance with its conflict of interest policies.

Abbreviations: APC, adenomatous polyposis coli; PE, phycoerythrin.

†To whom correspondence should be addressed. E-mail: vogelbe@jhmi.edu.

© 2005 by The National Academy of Sciences of the USA

MAY 21, 2002

and 98°C for 10 s, 61°C for 10 s, and 72°C for 10 s 30 times. Each reaction was performed in duplicate, and a calibration curve was generated in each 96-well plate by using various amounts of normal human genomic DNA. The concentration of PCR products was determined by using a PicoGreen dsDNA quantification assay (Invitrogen).

BEAMing. A common oligonucleotide (5'-TCCCGCGAAATTA-ATACGAC-3') was synthesized with a dual biotin group at the 5' end and with a six-carbon linker (C6) between the biotin and the other nucleotides (Integrated DNA Technologies, Coralville, IA). This oligonucleotide was coupled to streptavidin-coated magnetic beads (MyOne, Dynal, Oslo) according to the protocol described in ref. 16. The water-in-oil emulsions were prepared by modifications of the methods described by Ghadessy and Holliger (17) and Bernath *et al.* (18). For each emulsion PCR, a 240- μ l aliquot of an aqueous PCR mix was added to 960 μ l of 7% (wt/vol) Abil EM90 (Degussa Goldschmidt Chemical, Hopewell, VA) in mineral oil (Sigma). The aqueous phase contained 67 mM Tris-HCl (pH 8.8), 16.6 mM (NH₄)₂SO₄, 6.7 mM MgCl₂, 10 mM 2-mercaptoethanol, a 0.2 mM concentration of each dNTP, 0.05 μ M forward primer (5'-TCCCGCGAAATTAATACGAC-3'), 8 μ M reverse primer, 0.2 units/ μ l Platinum Taq polymerase (Invitrogen), 3 \times 10⁵ per μ l oligonucleotide-coupled beads, and 0.1 pg/ μ l template DNA. The reverse primers are listed in the supporting information. The water-oil mix was vortexed for 10 s and then emulsified for 50 s by using an Ultra-Turrax homogenizer (T25 basic, IKA, Wilmington, NC) with a disposable OmniTip (Omni International, Waterbury, CT) at the minimum speed. The emulsions were aliquoted into 8 wells of a 96-well PCR plate and cycled under the following conditions: 94°C for 2 min; 94°C for 10 s, 58°C for 15 s, and 70°C for 15 s 50 times. After PCR, the emulsions were pooled into a 15-ml tube and demulsified through the addition of 10 ml of NX buffer (100 mM NaCl/1% Triton X-100/10 mM Tris-HCl, pH 7.5/1 mM EDTA/1% SDS). After vortexing for 10 s, the beads were pelleted by centrifugation for 5 min at 4,100 \times *g*. The top phase was removed, and the beads were resuspended in 800 μ l of NX buffer and transferred to a 1.5-ml tube. The beads were collected by using a magnet (MPC-S, Dynal) and washed with 800 μ l of wash buffer (20 mM Tris-HCl, pH 8.4/50 mM KCl). The double-stranded DNA on the beads was converted to single-stranded DNA by incubation in 800 μ l of 0.1 M NaOH for 2 min at room temperature. The beads were washed twice with 800 μ l of wash buffer, using the magnet, and finally resuspended in 200 μ l of wash buffer. Single base extension and flow cytometry were performed as described in the supporting information.

Results

Circulating Mutant DNA Is Degraded. We used real-time PCR or digital PCR to determine the number of total circulating *APC* (adenomatous polyposis coli) genes in 33 patients with colorectal tumors and 10 age-matched donors without any tumors. The number of *APC* gene copies was significantly higher in advanced stage patients (Dukes' D) than in patients with early stage cancers ($P < 0.0001$, Student's *t* test), consistent with previous studies (19, 20). In advanced stage patients, the median number of *APC* gene fragments per ml of plasma was 47,800, whereas the median number was 3,500 and 4,000 for patients with Dukes' A and Dukes' B cancers, respectively (Table 1). There was no significant difference between the number of circulating copies in early stage cancer patients (Duke's A or B), patients with adenomas (4,300 *APC* fragments per ml of plasma), and normal individuals (3,460 *APC* fragments per ml of plasma; range of 1,150–8,280 fragments per ml).

To determine the size of mutant gene fragments in circulating DNA, we analyzed plasma DNA from three patients with advanced colorectal cancers (Dukes' D, metastatic to liver) who were shown to contain *APC* gene mutations in their tumors. By varying the size

of the amplicons, it was possible to determine the number of normal and mutant gene fragments by sequencing the PCR products derived from one or a few template molecules (detailed in the supporting information). The size of the amplicons varied from 100 to 1,296 bp and encompassed the mutation present in each patient. The number of total *APC* fragments (WT plus mutant) increased by 5- to 20-fold as the size of the amplicons decreased from 1,296 to 100 bp (Fig. 1A). The fraction of mutant molecules was strikingly dependent on size of the amplicon, increasing by >100-fold over the size range tested (Fig. 1B).

We conclude that the mutant DNA fragments present in the circulation of cancer patients are degraded compared with the circulating DNA derived from nonneoplastic cells. This conclusion is consistent with previous studies of other tumor types (21, 22) and has important implications for the detection of such mutant molecules.

Development of a Quantitative Assay for Detection of Rare Mutations.

The results described above were obtained by sequencing hundreds of PCR products, each derived from one or a few DNA template molecules. In preliminary studies, we found that such digital PCR-based techniques were sufficiently sensitive to detect circulating mutant DNA molecules in patients with advanced cancers but not in patients with early stage cancers. To increase the sensitivity and reliability of these assays, we developed an extension of BEAMing (which derives its name from its principal components: beads, emulsion, amplification, and magnetics) that allowed us to examine many more template molecules in a convenient fashion. The approach consists of four steps. (i) Real-time PCR was used to determine the number of total *APC* gene fragments in the plasma sample (Fig. 2A, step 1). (ii) BEAMing was used to convert the amplified plasma DNA into a population of beads (Fig. 2A, steps 2–4). (iii) The mutational status of the extended beads was determined by single base extension (Fig. 2B). (iv) Flow cytometry was used to simultaneously measure the FITC, Cy5, and phycoerythrin (PE) signals of individual beads.

Fig. 3 shows a representative flow cytometry result wherein the interpretation of the profiles was confirmed experimentally. In the example shown, 342,573 beads were analyzed by flow cytometry. The single bead population (295,645) was used for the fluorescence analysis (Fig. 3A). Of these, 30,236 exhibited a PE signal (Fig. 3B), indicating that they had been extended during the emulsion PCR. The FITC and Cy5 signals reflected the number of beads containing mutant or WT sequences, respectively. Beads containing the WT DNA sequences (30,186) had high Cy5 but background FITC signal ("red beads" in Fig. 3C). Beads extended only with mutant DNA sequences (22) had high FITC signals but background Cy5 signals ("green beads"). Twenty-eight had both FITC and Cy5 signals ("blue beads"). Such dual-labeled beads resulted from either the presence of both a WT and mutant template in the droplet containing the bead or an error in the early cycles of the emulsion PCR (see below). These dual-labeled beads were eliminated from analysis, and only homogeneously labeled beads were considered for the enumeration of mutations. Note that this conservative analysis strategy results in a slight underestimation of the fraction of mutations, because it excludes mutants that were present in droplets that also contained one or more WT fragments. Beads in each of these three populations were collected by flow sorting, and single beads from the sort were used as templates in conventional DNA sequencing. All 131 beads subjected to sequencing analysis showed the expected patterns, with examples illustrated in Fig. 3C.

Limits to the Sensitivity of Assays for Plasma DNA Mutations. The results described above show that the BEAMing approach can, in principle, detect a very small fraction of fragments containing mutant sequences within a much larger pool of fragments containing WT sequence. Because >50 million beads are used in a single emulsion PCR, and flow cytometry can be performed

Table 1. Quantification of APC mutations in plasma

Patient no.	Sex/ age, yr	Site	Dukes' stage (tumor node metastasis stage)	Diameter of lesion, cm	Mutation identified in primary tumor (codon)	Fragments per ml of plasma	No. of fragments analyzed	Percentage of mutant fragments, %
1	M/50	Ascending colon	Adenoma	3.0	C4348T (1450)	2,600	2,350	0.002
2	M/67	Descending colon	Adenoma	2.5	C4285T (1429)	5,080	5,080	0.001
3	M/54	Rectum	Adenoma	4.0	G3856T (1286)	4,150	4,150	0.002
4	F/82	Rectum	Adenoma	3.0	4147-4148insA (1383)	1,350	1,350	0.001
5	F/65	Rectum	Adenoma	1.0	C4067G (1356)	4,260	4,260	0.001
6	F/71	Ascending colon	Adenoma	4.0	G3856T (1286)	4,150	4,150	0.001
7	M/68	Cecum	Adenoma	6.5	C4285T (1429)	4,760	4,760	0.003
8	M/93	Ascending colon	Adenoma	0.8	A4345T (1449)	4,320	4,320	0.001
9	F/78	Ascending colon	Adenoma	3.0	C4216T (1406)	28,570	28,570	0.001
10	F/59	Sigmoid colon	Adenoma	5.0	4661-4662insA (1554)	2,160	2,160	0.002
11	F/73	Ascending colon	Adenoma	5.0	C4348T (1450)	8,000	8,000	0.02
Median/mean						4,300/6,300		0.02*
Mutant plasma samples per samples analyzed								1/11 (9)
12	F/81	Sigmoid colon	A (T2N0M0)	4.0	G4189T (1397)	7,900	12,000	0.01
13	F/75	Sigmoid colon	A (T2N0M0)	2.5	3927-3931del AAAGA (1309)	2,160	2,160	0.001
14	M/60	Sigmoid colon	A (T2N0M0)	3.0	3927-3931del AAAGA (1309)	4,600	6,900	0.04
15	M/79	Right colic flexure	A (T2N0M0)	3.0	4470delT (1490)	4,600	3,696	0.03
16	M/70	Ileocecal	A (T2N0M0)	2.5	4481delA (1494)	6,200	3,105	0.07
17	F/68	Ascending colon	A (T2N0M0)	3.5	C4348T (1450)	2,170	2,170	0.001
18	F/66	Sigmoid colon	A (T1N0M0)	2.5	3927-3931del AAAGA (1309)	1,920	1,920	0.001
19	M/68	Rectum	A (T2N0M0)	5.5	C3907T (1303)	2,300	1,170	0.12
Median/mean						3,500/4,000		0.04/0.04*
Mutant plasma samples per samples analyzed								5/8 (63)
20	F/65	Cecum	B (T3N0M0)	3.5	G4396T (1466)	5,300	5,300	0.002
21	M/71	Sigmoid colon	B (T3N0M0)	3.0	C4348T (1450)	2,100	1,863	0.19
22	M/37	Descending colon	B (T4N0M0)	10.0	C4330T (1444)	5,400	4,887	1.28
23	M/64	Sigmoid colon	B (T3N0M0)	6.5	C4099T (1367)	3,810	3,810	0.001
24	M/72	Sigmoid colon	B (T3N0M0)	3.0	C4012T (1338)	4,800	4,800	0.03
25	F/82	Hepatic flexure	B (T3N0M0)	4.0	C4099T (1367)	3,840	3,840	1.46
26	M/83	Ascending colon	B (T3N0M0)	6.0	4470delT (1490)	1,600	1,404	1.75
27	M/61	Sigmoid colon	B (T3N0M0)	4.0	4260-4261delCA (1420)	4,200	4,200	0.001
Median/mean						4,000/3,900		1.28/0.94*
Mutant plasma samples per samples analyzed								5/8 (63)
28	F/83	Ascending colon	D (T3N2M1)	5.0	4661-4662insA (1554)	230,000	24,857	5.6
29	M/55	Sigmoid colon	D (T3N0M1)	3.0	G3925T (1309)	69,600	1,636	27.4
30	F/33	Descending colon	D (T4N1M1)	5.0	C4067A (1356)	18,000	491	10.5
31	M/64	Sigmoid colon	D (T4N2M1)	6.0	T4161A (1387)	26,000	975	1.9
32	M/56	Rectum	D (T3N2M1)	3.0	4468-4469delCA (1490)	103,200	1,187	18.9
33	F/60	Rectum	D (T3N2M1)	4.0	4059-4060insT (1354)	8,400	850	2.0
Median/mean						47,800/75,900		8.05/11.05*
Mutant plasma samples per samples analyzed								6/6 (100)

*Calculated only for samples in which the percentage of mutant fragments was significantly higher than in control samples (i.e., >0.003%; printed in boldface). M, male; F, female.

at speeds of >50,000 beads per s, the capacity to enumerate such mutations is not limited by the beads themselves. Instead, two other features limit the sensitivity. First, there is a finite number of DNA fragments present in clinical samples. As noted above, this number ranged from 1,350 to 230,000 fragments per ml in the patients with tumors (Table 1) and from 1,150 to 8,280 fragments per ml in control patients, which gives an upper bound to the sensitivity of the assays. For example, a calculation using the Poisson distribution shows that if 4,000 fragments were

analyzed, the mutation fraction in circulating DNA would have to be >1 in 1,333 fragments (i.e., 3 divided by the number of total fragments analyzed) for the assay to achieve 95% sensitivity. A second limiting feature is the error rates of the polymerases used for PCR. In our approach, two PCR steps are used: The first is a conventional PCR that employs plasma DNA fragments as templates, and the second is an oil-in-water emulsion PCR that uses the initial PCR products as templates. In the emulsion PCR, errors occurring during the early rounds of PCR can result in

MAY 21, 2022




Fig. 1. Effect of the PCR amplicon size on plasma DNA concentration and mutation frequency. (A) The concentration of total *APC* fragments (WT plus mutant) of various sizes was determined by using digital PCR of plasma DNA from three different patients (patients 29, 30, and 32). (B) The fraction of mutant *APC* fragments was determined by digital sequencing of PCR products.

heterogeneous beads containing both WT and mutant sequences. These beads are easily eliminated from consideration, as described in Fig. 3C. However, the errors introduced in the first PCR cannot be eliminated, because they give rise to beads with homogeneous mutant sequences, indistinguishable from those resulting from genuine mutations in the original plasma DNA templates.

The fraction of mutant molecules present after the first PCR equals the product of the mutation rate of the polymerase and the number of cycles carried out. BEAMing provides a quantitative way to determine the error rate of any polymerase used in PCR without requiring cloning in bacterial vectors (M.L., F.D., S.N.G., K.W.K., and B.V., unpublished data). Of 19 different base changes evaluated in normal DNA, the error rates with the polymerase used in the current study averaged 3.0×10^{-7} mutations per bp per PCR cycle and ranged from 1.7×10^{-7} to 6.5×10^{-7} mutations per bp per PCR cycle, depending on the mutation site assessed. As a result, we only scored plasma samples as positive for mutations if their frequency in the sample was significantly higher than the maximum error rate of polymerase found experimentally (i.e., 1.95×10^{-5} after 30 cycles). As a result of the relatively low error rate with the polymerase used, it was the number of molecules present in the original plasma sample, rather than the polymerase error rate *per se*, that limited sensitivity.

These issues suggest that the sensitivity of assays for circulating mutant DNA could be increased in the future by (i) the development of new or modified polymerases with reduced error rates and (ii) the use of more plasma per assay (i.e., more template molecules).




Fig. 3. Processing of flow cytometry data obtained by BEAMing. (A) Dot plot of forward-scatter (FSC) and side-scatter (SSC) signals of beads. (B) Histogram of single beads with regard to PE signal. (C) Dot plot showing the Cy5 and FITC fluorescence intensity profiles of PE-positive beads. The beads clustered in three distinct populations colored red, green, and blue. Sequencing of individual beads sorted from each population showed that the red and green beads contained homogeneous WT and mutant sequences, respectively; the blue beads contained a mixture of WT and mutant sequences.

Quantification of Mutant *APC* Fragments in Plasma from Patients with Colorectal Tumors. Based on the principles derived from the experiments described above, we determined whether fragments of tumor DNA could be detected in patients with colorectal tumors of various types. We selected *APC* gene mutations for this assessment, because >85% of colorectal tumors contain mutations of this gene, irrespective of tumor stage (23). Mutations within codon 1209–1581 of *APC*, containing most previously identified mutations, were evaluated by sequencing of DNA purified from the tumors of 56 patients. Mutations were observed in 33 of these patients (59%), and, as expected, the proportion of tumors with these mutations did not differ significantly among tumors of various stages (see the supporting information).

A BEAMing assay was then designed for each of the mutations identified in the 33 tumors and applied to the DNA purified from the plasma of the corresponding patients (Table 1). In each case, DNA from normal lymphocytes or plasma from patients without




Fig. 2. Schematic of the BEAMing-based assay. (A) Extended beads were prepared by modifications of the BEAMing procedure described by Dressman *et al.* (16). (B) Single base extensions were performed on the extended beads. Normal DNA sequences contained a G at the queried position; mutant sequences contained an A.




Fig. 4. Examples of flow cytometric profiles of beads generated from plasma DNA (patient 16). Cy5 and FITC fluorescence intensity profiles of PE-positive beads from four patients are shown. The patients, mutations, and fraction of mutant *APC* fragments are indicated.

cancer was used as a negative control. DNA from the tumors of the 33 patients was used as a positive control. All six patients with advanced lesions (Dukes' D, defined as having at least one distant metastatic lesion) were found to contain mutant DNA fragments in their plasma. Among 16 patients harboring cancers with a favorable prognosis (Dukes' A or B, defined as having no lymph node involvement and no distant metastases), 10 (63%) were found to contain mutant DNA fragments in their plasma. In contrast, among 11 patients with large, benign tumors (adenomas), only 1 patient's plasma was found to contain mutant DNA fragments. Representative flow cytometric results are shown in Fig. 4 and summarized in Table 1.

The fraction of mutant molecules found in the plasma of the 17 cases with detectable mutations also varied according to tumor stage ($P < 0.0001$, Fisher exact test). In the advanced cases (Dukes' D), an average of 11.1% (range of 1.9–27%) of the total *APC* gene fragments were mutant. In patients without metastases (Dukes' B), an average of 0.9% (range of 0.03–1.75%) of the plasma *APC* gene fragments were mutant. In patients with lower stage tumors (Dukes' A), the fraction was even lower, averaging 0.04% (range of 0.01–0.12%). And in the one patient with a benign tumor, only 0.02% of the plasma DNA fragments were mutant. The median fraction of positive beads found in the control DNA samples from patients without cancer was 0.0009% (range of 0.003–0.0005%). The mutations in the control samples likely resulted from PCR errors, as noted above.

Table 1 also lists the concentration of total *APC* fragments (WT plus mutant) in these patients' plasma. There was no direct relationship between the concentration of total *APC* fragments and the mutational load. Although patients with advanced cancers tended to have higher concentrations of total *APC* fragments than the other patients, this increase was not due to DNA from neoplastic cells. Furthermore, no correlation was found between tumor burden (volume of primary tumor plus metastatic sites) and either the concentration of *APC* fragments or percentage of mutant *APC* fragments in the circulation.

Discussion

The data described above conclusively demonstrate that *APC* gene fragments from the neoplastic cells of colorectal tumors




Fig. 5. Fraction of mutant *APC* gene fragments in the plasma of patients with various colorectal tumors [adenomas (Ad) and Dukes' stage A, B, and D carcinomas]. In each mutation analyzed, DNA from normal lymphoid cells or plasma DNA from healthy donors was used as a control (Normal). The "mutants" observed in assays with normal cellular DNA represent errors generated during the PCR process rather than mutations present in the template DNA (see text). The red lines represent the mean, minimum, and maximum values of the normal controls.

can be found in the circulation and that the number of such fragments depends on tumor stage. These results have implications for both colorectal tumor biology and for practical diagnostic tests, as discussed below.

Source of Plasma DNA. Previous studies have shown that the total DNA concentration in the plasma of cancer patients is often elevated (19, 20). Our results support this conclusion only in advanced stage patients, in that more total *APC* gene fragments (WT plus mutant) were present in the plasma of patients with Dukes' D cancers than in those with earlier stage tumors. Our results additionally show that this "extra" DNA in advanced stage patients is not derived from the neoplastic cells themselves, because only a minor fraction of the circulating *APC* fragments are mutant, whereas all of the neoplastic cell's *APC* fragments are mutant.

But there are still a large number of mutant DNA fragments circulating in cancer patients. Assuming that the volume of distribution of DNA at steady state is similar to that of oligonucleotides in primates (60–70 ml/kg), an 8% fraction of mutant molecules among 47,800 fragments per ml of plasma (as in Dukes' D patients) would correspond to 1.6×10^7 mutant fragments present in a 70-kg person at any given time (24). The half-life of this tumor DNA is estimated at 16 min, based on the data obtained from clearance of fetal DNA in maternal plasma (25), which translates to $\approx 6 \times 10^8$ mutant fragments released from the tumor each day. For patients with a tumor load 100 g in size ($\approx 3 \times 10^{10}$ neoplastic cells), we thereby estimate that 3.3% of the tumor DNA is fed into the circulation on a daily basis. For a Dukes' B cancer of 30 g in which 1.3% of the 4,000 circulating *APC* fragments per ml of plasma are mutant, the corresponding estimate is that 0.15% of the tumor DNA is fed into the circulation each day.

So how do mutant *APC* gene fragments get into the plasma? Several clues are provided by our data. The ability to get into the circulation was clearly not related to tumor size, because the benign tumors we studied were as large as the cancers (Table 1), yet the former rarely gave rise to detectable mutant DNA fragments. Similarly, there was no significant correlation between the tumor load (including metastatic deposits) and the amount of mutant DNA in the circulation. In contrast, the degree of invasion was indeed correlated with the number of circulating DNA fragments. Those lesions that weren't invasive (benign tumors) did not commonly feed mutant DNA molecules into the plasma. As tumors

Explore Litigation Insights

Docket Alarm provides insights to develop a more informed litigation strategy and the peace of mind of knowing you're on top of things.

Real-Time Litigation Alerts



Keep your litigation team up-to-date with **real-time alerts** and advanced team management tools built for the enterprise, all while greatly reducing PACER spend.

Our comprehensive service means we can handle Federal, State, and Administrative courts across the country.

Advanced Docket Research



With over 230 million records, Docket Alarm's cloud-native docket research platform finds what other services can't. Coverage includes Federal, State, plus PTAB, TTAB, ITC and NLRB decisions, all in one place.

Identify arguments that have been successful in the past with full text, pinpoint searching. Link to case law cited within any court document via Fastcase.

Analytics At Your Fingertips



Learn what happened the last time a particular judge, opposing counsel or company faced cases similar to yours.

Advanced out-of-the-box PTAB and TTAB analytics are always at your fingertips.

API

Docket Alarm offers a powerful API (application programming interface) to developers that want to integrate case filings into their apps.

LAW FIRMS

Build custom dashboards for your attorneys and clients with live data direct from the court.

Automate many repetitive legal tasks like conflict checks, document management, and marketing.

FINANCIAL INSTITUTIONS

Litigation and bankruptcy checks for companies and debtors.

E-DISCOVERY AND LEGAL VENDORS

Sync your system to PACER to automate legal marketing.